



> ICE-100 Incubation Chamber




> Product Specification

Time Setting	3-30 mins
Temperature Setting	20°-40°C
Humidity	≤95%
Atmosphere Pressure	86.0-106.0KPa
Operating Altitude	<2000m
Power Input	DC12V, 5.0A
Dimension	220x160x72mm
Weight	1.4kg


> Test Procedure




Turn on the power switch



Press the SET button and adjust the temperature setting



Press the SET button and adjust the time setting



Insert the reagent card and start to incubate